

Data Collection Statistics
Montezuma County - Colorado

Preliminary Validation
06/01/2012 - 06/30/2012

Parameter	Interval	Par Code	Data Recovery			Valid Data	
			No. Possible	No. Collected	% Collected	No. Valid	% Valid
Ozone	hourly	O3	720	712	98.9	712	98.9
Station Temperature	hourly	STP	720	717	99.6	717	99.6

Note: The percent valid is calculated against the number possible.

Automatic zeros and spans are performed daily on most ambient gas analyzers, therefore, no ambient data can be collected during this time. As a result, the maximum percent valid for ambient gas data typically can not be greater than 95.8.

Performance Goals:


Quarterly Criteria:

100% of sites, >= 85% valid data capture
90% of sites, >= 90% valid data capture
80% of sites, >= 95% valid data capture


Monthly Criteria:

100% of sites, >= 60% valid data capture
90% of sites, >= 75% valid data capture
80% of sites, >= 85% valid data capture

Montezuma County, Colorado - Cortez County Health Building




Montezuma County, Colorado - Cortez County Health Building




June 2012

Montezuma County, Colorado - Cortez County Health Building



Montezuma County, Colorado - Cortez County Health Building



Ozone
 Summary of the Highest Daily Maximum 8-Hour Averages
 Montezuma County - Colorado

Preliminary Validation
 06/01/2012 - 06/30/2012

Date	Start and End Time of Daily Maximum 8-Hour Average	Daily Maximum 8-Hour Average (ppb)
06/20/2012	11 - 18	67
06/21/2012	09 - 16	66
06/03/2012	10 - 17	65
06/10/2012	10 - 17	63
06/24/2012	11 - 18	62
06/04/2012	10 - 17	60
06/23/2012	10 - 17	60
06/19/2012	11 - 18	58
06/25/2012	09 - 16	58
06/16/2012	10 - 17	57
0	Days this period with 8-hour averages > .075 ppm (75 ppb)	